

CERTIFICATE OF ANALYSIS

ERM[®]-CC015a

ausverkauft / out of stock

Mineral oil contaminated sediment		
	Certified value ¹⁾	Uncertainty ²⁾
Compound	Mass fraction in mg/kg	
Total petrol hydrocarbons (TPH)	1820	± 130
¹⁾ Unweighted mean value of 11 laboratory means using gas chromatography with flame ionisation detection (GC/FID) according to ISO/FDIS 16703:2003.		
²⁾ Estimated expanded uncertainty <i>U</i> with a coverage factor of <i>k</i> =2, corresponding to a level of confidence of 95 %, as defined in the Guide to the Expression of Uncertainty in Measurement, ISO, 1993.		

This certificate is valid for a period of 12 months beginning with the dispatch of the reference material from BAM.

The minimum sample size for one determination is 10 g. The mass fraction of TPH is related on sample intake (not on dry mass). The water content is 5.7 % (*Karl Fischer* titration) and remains stable if the material is handled as indicated below.

NOTE

European Reference Material ERM[®]-CC015a replaces ERM[®]-CC015 which is sold out. It was produced and certified under the responsibility of Bundesanstalt für Materialforschung und –prüfung (BAM) according to the principles laid down in the technical guidelines of the European Reference Materials[®] co-operation agreement between BAM-LGC-IRMM. Information on these guidelines is available on the Internet (<http://www.erm-crm.org>).

Accepted as an ERM[®], Berlin, 2005-06-08

Berlin, 2005-06.08

BAM Berlin
Department I
Analytical Chemistry;
Reference Materials
12200 Berlin, Germany

BAM Berlin
Division I.2
Organic Chemical Analysis;
Reference Materials
12200 Berlin, Germany

Prof. Dr. U. Panne
(Head of Department)

Prof. Dr. I. Nehls
(Head of Division)

DESCRIPTION OF THE SAMPLE

The intended purpose of reference material ERM[®]-CC015a is the verification of analytical procedures for the determination of TPH in soils and sediments according to ISO/FDIS 16703:2003 [1] by GC/FID and for quality control in analytical laboratories.

The material ERM[®]-CC015a is a sediment sampled from River Weisse Elster, near the city of Leipzig, Saxony, Germany. The material had been contaminated over decades by industrial and municipal sewages. After drying, fractionation by sieving and homogenisation, the sieving fraction < 125 µm was subdivided into 360 units of 81 g which were filled in brown glass bottles with screw caps equipped with PTFE-inserts and sealed with shrinking foil. The material is stored at BAM at –20°C until dispatch.

The within- and among-bottle-homogeneity of the TPH content was demonstrated by analysis of variance on 15 selected bottles (5 replicate analyses per bottle).-

The initial stability study after storage of selected units of this reference material for up to 6 months at different temperatures allows to guarantee the period of validity of the certificate. Periodical investigations on the stability of this material will be carried out in order to keep this information up to date. The tests for homogeneity and stability are described in detail in the technical report.

PARTICIPANTS

The following laboratories, including three independent operator/equipment combinations at BAM, participated in the certification study using GC/FID for the determination of TPH according to ISO/FDIS 16073:2003 [1]:

chemlab, Gesellschaft für Analytik und Umweltberatung mbH	64625	Bensheim
Umweltbundesamt, Labor für Wasseranalytik, FG II 2.5	14193	Berlin
Bundesanstalt für Materialforschung und -prüfung (BAM), Division I.2	12489	Berlin
CDM Amann Infutec Consult AG & Co. KG	74564	Crailsheim
Handels- und Umweltschutzlabor Dr. Kaiser & Dr. Woldmann GmbH	22761	Hamburg
Behörde für Umwelt und Gesundheit	20539	Hamburg
Dr. Ronald Fischer AÜB, Chemische Analytik- und Umweltberatung	99894	Leinatal
Laborgesellschaft für Umweltschutz m. b. H.	67433	Neustadt/Weinstrasse
PROTEKUM Umweltinstitut GmbH Oranienburg	16515	Oranienburg
Limnologisches Institut Dr. Karl Ernst Nowak	28870	Ottersberg

SAFETY INFORMATION

It is strongly recommended to handle and dispose the reference material in accordance with the guidelines for hazardous materials legally in force at the site of end use and disposal.

INSTRUCTIONS FOR USE AND STORAGE

The material has to be stored at -20°C in its original bottle. Before withdrawing a subsample the bottle has to have reached ambient temperature. Thereafter, the bottle is to be closed tightly and stored at –20°C. The stability of the reference material is not affected by short periods of handling at ambient temperature during transport and use.

TECHNICAL REPORT

A detailed technical report (paper copy) describing the production, general characterisation as well as the analytical procedures applied and the treatment of the analytical data during certification of ERM[®]-CC015a is available on request from BAM.

LEGAL NOTICE

Neither the Federal Institute for Materials Research and Testing (BAM) nor any person acting on their behalf make any warranty or representation, express or implied, that the use of any information, material, apparatus, method or process disclosed in this document may not infringe privately owned rights, or assume any liability with respect to the use of, or damages resulting from the use of any information, material, apparatus, method or process disclosed in this document.

REFERENCES

- [1] ISO /FDIS 16703:2003, "Soil quality - Determination of mineral oil content by gas chromatography"

Supply of Reference Materials by Bundesanstalt für Materialforschung und –prüfung:

Richard-Willstätter-Straße 11, 12489 Berlin, Germany

Phone: +49 30 8104 5821

e-mail: sales.crm@bam.de

Fax: +49 30 8104 1117

internet: www.bam.de